

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

JUL 09 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130287</u>			
Supervisor/Operator Performing the Verification Procedure:			
Name	<u>Officer Kelly George</u>	ID#	<u>4447</u> Date <u>7-4-09</u> ✓
A Agency	<u>Kenai Police Dep.</u>	Phone #	
Instrument Location	<u>Wildwood Pretrial Facility, Kenai</u> ✓		
B Alco S/N	<u>X301542</u>	Target Value	<u>.082</u> High Pressure <u>550</u>
Alco Test Values	<u>.083</u>	<u>.082</u>	
	1 st Alco	2 nd Alco	
Signature	<u>Kelly George</u>		<u>CB</u> <u>7/14/09</u>

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 21st day of July, 2009.

Carolyn M. Noland
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130287 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130287

JULY 04, 2009

OPERATOR'S NAME:

GEORGE/KELLY

OPERATOR'S NUMBER: 4447

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI:

T

O.L. #: T

DEPT/AGENCY: KENI

CASE/REPORT:

TEST TYPE: U

ALCO TARGET VALUE: .082 ✓

ALCO S/N: X301542 ✓

--- BREATH ANALYSIS ---

.082 ADJUSTED FOR 30.37 in
ALCO TARGET .083 02:09
BLANK TEST .000 02:10
INTERNAL STANDARD VERIFIED 02:10
ALCO TU 30.37 in .083 ✓ 02:11
BLANK TEST .000 02:11
SUBJECT SAMPLE *COB 7/14/09* .080 02:12
BLANK TEST .000 ✓ 02:13
ALCO TU 30.37 in .082 ✓ 02:13
BLANK TEST .000 02:14

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130287

JULY 04, 2009 ✓

TIME 02:05

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 42c
BAROMETER: 30.37 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!##%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~