

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

OCT 22 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130287</u>
A	Name	<u>Mitch Langseth</u>	ID# <u>2612</u> Date <u>10-8-09</u>
	Agency	<u>KENAI PD</u>	Phone # <u>283-7879</u>
	Instrument Location	<u>Wildwood Pretrial</u>	
B	Alco S/N	<u>X301765</u>	Target Value <u>.085</u> High Pressure <u>800</u>
	Alco Test Value Average	<u>.084</u> ✓	<u>.085</u> ✓
		1 st Alco	2 nd Alco
	Signature		<u>COS</u> <u>10/27/09</u>

(OVER)


(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 6th day of Nov., 2009.


(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130287

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130287

OCTOBER 08, 2009

OPERATOR'S NAME:

LANGSETH

OPERATOR'S NUMBER: 2612

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

Z

O.I.L. #: 0

DEPT/AGENCY: KENI

CASE/REPORT: 0

TEST TYPE: U

ALCO TARGET VALUE: .085

ALCO S/N: K301785

--- BREATH ANALYSIS ---

.085 ADJUSTED FOR 30.87 in
ALCO TARGET .085 06:45
BLANK TEST .000 06:47
INTERNAL STANDARD VERIFIED 06:47
ALCO TV 30.87 in .084 06:47
BLANK TEST .000 06:48
SUBJECT SAMPLE .000 06:48
BLANK TEST .000 06:49
ALCO TV 30.87 in .085 06:50
BLANK TEST .000 06:50

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130287

OCTOBER 08, 2009

TIME 06:51

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 42c

BAROMETER: 30.87 in

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!##%&'()*+,-./0123456789:;<=>@ABCDEFGHIJK
LMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~■