

JAN 03 2010 ²⁰¹¹ 3/4/11

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: **DataMaster cdm S/N 130287** ✓

Name Kelly George ID# 4447 Date 12-28-10 ✓

A Agency Kenai Police Department Phone # 283-7879

Instrument Location Wildwood Pretrial Facility, Kenai

B Alco S/N X301600 ✓ Target Value .082 ✓ High Pressure 1100

Alco Test Values .079 ✓ 1st Alco .079 ✓ 2nd Alco


Signature Kelly George COB
1/8/11

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.




Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 7th day of MAR, 2011

 (Notary Seal Stamp)

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130287

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130287

DECEMBER 28, 2010

OPERATOR'S NAME:

GEORGE KELLY

OPERATOR'S NUMBER: 4447

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

UOC

O.L. #: UOC

DEPT/AGENCY: KEN1

CASE/REPORT: UOC

TEST TYPE: U

ALCO TARGET VALUE: .082

ALCO S/N: X301600

--- BREATH ANALYSIS ---

.082 ADJUSTED FOR 29.60 in
ALCO TARGET .081 14:00
BLANK TEST .000 14:01
INTERNAL STANDARD VERIFIED 14:01
ALCO TV 29.60 in .079 14:02
BLANK TEST .000 14:02
SUBJECT SAMPLE .000 14:03
BLANK TEST .000 14:04
ALCO TV 29.60 in .079 14:04
BLANK TEST .000 14:05

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130287

DECEMBER 28, 2010

TIME 13:56

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 42c

BAROMETER: 29.60 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~