# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument*
*State of Alaska*
*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

**DataMaster cdm S/N** 130288

Supervisor/Operator Performing the Verification Procedure:

A

Name DIXIE SPENCER ID# 2280 Date 1/27/08

Agency AST Phone # 675-4398

Instrument Location ANIAK

B

Alco S/N X172931 Target Value .077 High Pressure 710

Alco Test Values .077 - 8 (1st Alco) .077 - 8 02/10/08 (2nd Alco)

Signature Dixie Spencer 02/19/08

(OVER)

Under the Alaska Rules of Evidence, I certify that: (Do Not Print in the area below)

(a) I am Jeanne Swartz , Criminalist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Dixie Spencer , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

02/22/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORIES

BT9 03/00

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm breath Test Instrument*
*State of Alaska*
*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

**DataMaster cdm S/N** 130288

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130288

JANUARY 27, 2008 CDB 2/19/08

OPERATOR'S NAME:
SPENCER DIXIE W
OPERATOR'S NUMBER: 2280
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
Z
O.L. #: Z
DEPT/AGENCY: AN10
CASE/REPORT: 08
TEST TYPE: V
ALCO TARGET VALUE: .077
ALCO S/N: X172931

--- BREATH ANALYSIS ---

.077 ADJUSTED FOR 29.54 in

| | | |
|---|---|---|
| ALCO TARGET | .076 | 11:58 |
| BLANK TEST | .000 | 12:00 |
| INTERNAL STANDARD | VERIFIED | 12:00 |
| ALCO TV 29.54 in | .077 | 12:00 |
| BLANK TEST | .000 | 12:01 |
| SUBJECT SAMPLE | .000 | 12:02 |
| BLANK TEST | .000 | 12:02 |
| ALCO TV 29.54 in | .077 | 12:03 |
| BLANK TEST | .000 | 12:03 |

CDB 2/19/08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130288

JANUARY 27, 2008
TIME 12:26

--- DIAGNOSTIC CHECK ---

| | |
|---|---|
| COMPUTER: | OKAY |
| PROGRAM: | OKAY |
| SOFTWARE DATE: | 02/20/01 |
| HEATERS | |
| SAMPLE CHAMBER: | 49c |
| BREATH TUBE: | 42c |
| BAROMETER: | 29.52 in |
| FLOW DETECTOR: | OKAY |
| PUMP | |
| HIGH SPEED: | OKAY |
| DETECTOR: | OKAY |
| FILTERS: | OKAY |
| QUARTZ STANDARD: | OKAY |
| CALIBRATION: | OKAY |

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~