

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130288</u> ✓
Name	<u>DIXIE SPENCER</u>	ID# <u>2280</u> Date <u>2/8/09</u> ✓
A Agency	<u>AST</u>	Phone # <u>475-4398</u>
Instrument Location	<u>ANIAC</u> ✓	
Alco S/N <u>X301605</u> ✓ Target Value <u>.079</u> ✓ High Pressure _____		
B Alco Test Values	<u>.079</u> ✓ 1 st Alco 	<u>.084</u> ✓ 2 nd Alco
Signature	<u>Dixie Spencer</u> ✓	<u>CSB</u> <u>2/17/09</u>

(OVER)


Under the Alaska Rules of Evidence, (Do Not write in the area below)
I certify that,

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Dixie Spencer, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Juatz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

02/12/09

Date

BT9 03/00 FEB 10 2009

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130288

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130288

FEBRUARY 08, 2009 ✓ *COB 2/17/09*

OPERATOR'S NAME:
SPENCER DIXIE W
OPERATOR'S NUMBER: 2280
SUBJECT'S LAST NAME:
UNKNX XXXXX
SUBJECT'S FIRST NAME/MI :
X0000

D.L. #: 1234
DEPT/AGENCY: ANID
CASE/REPORT: 09
TEST TYPE: U
ALCO TARGET VALUE: .079 ✓
ALCO S/N: X301685

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.04 in		
ALCO TARGET	.079	09:39
BLANK TEST	.000	09:40
INTERNAL STANDARD	VERIFIED	09:40
3.00 TV 30.04 in	.085 ✓	09:40
BLANK TEST	.000	09:41
SUBJECT SAMPLE	.000	09:42
BLANK TEST	.000	09:42
3.00 TV 30.06 in	.084 ✓	09:43
BLANK TEST	.000	09:44

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130288

FEBRUARY 08, 2009 ✓
TIME 09:45

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	40c
BAROMETER:	30.06 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLNPQRSTUWVXYZ[\]^_`abcde fghijklmno
pqrstuvwxy{|}~!@