

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUN 16 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130288</u>
A	Name <u>DIXIE SPENCER</u>	ID# <u>2280</u> Date <u>6/15/09</u>
	Agency <u>AST ANIAK</u>	Phone # <u>625-4398</u>
	Instrument Location <u>ANIAK</u>	
B	Alco S/N <u>56799</u>	Target Value <u>.084</u> High Pressure <u>1100</u>
	Alco Test Values	
	<u>.085</u> 1 st Alco	<u>.085</u> 2 nd Alco
	Signature <u>Dixie Spencer</u>	
	(OVER)	<u>088</u> <u>6/16/09</u> <u>§ 6/16/09</u>

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 24th day of June, 2009.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130288 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130288

JUNE 15, 2009

OPERATOR'S NAME: *COB 6/14/09*
SPENCER DIXIE E
OPERATOR'S NUMBER: 2290
SUBJECT'S LAST NAME:
SUBJECT'S FIRST NAME/MI: 2222

O.L. #: 12345
DEPT/AGENCY: AN10
CASE/REPORT: 09000000
TEST TYPE: T
ALCO TARGET VALUE: .084 ✓
ALCO S/N: 56799 ✓

--- BREATH ANALYSIS ---

.084 ADJUSTED FOR 29.93 in		
ALCO TARGET	.084	10:31
BLANK TEST	.000	10:32
INTERNAL STANDARD	VERIFIED	10:32
ALCO TV 29.93 in	.085 ✓	10:33
BLANK TEST	.000	10:33
SUBJECT SAMPLE	.000	10:34
BLANK TEST	.000	10:35
ALCO TV 29.93 in	.085 ✓	10:35
BLANK TEST	.000	10:36

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130288

JUNE 15, 2009 ✓
TIME 10:37

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/28/01
FILTERS:	
SAMPLE CHANGER:	4c
BREATH TUBE:	4c
BAROMETER:	29.93 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
DIAGNOSTIC STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcedefghijklmnop
qrstuvwxyz{|}~