VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130288

Supervisor/Operator Performing the Verification Procedure:

A

Name DIXIE SPENCER ID# 2280 Date 12/14/08

Agency AST Phone # 675-4598

Instrument Location ANIAK

B

Alco S/N X301605 Target Value .079 High Pressure 1100

.086 8/ 12/20/08

Alco Test Values ~~.086~~ (1st Alco) .086 (2nd Alco)

Signature [signature] CDB 12/18/08

(OVER)

Under the Alaska Rules of Evidence, I certify that (Do Not Write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Dixie Spencer, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed

[signature] 12/20/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

BT9 03/00

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130288

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130288

DECEMBER 14, 2008

OPERATOR'S NAME: CDB 12/18/08
SPENCER DIXIE W
OPERATOR'S NUMBER: 2200
SUBJECT'S LAST NAME:
AAAA
SUBJECT'S FIRST NAME/MI :
AA
O.L. #: 12345
DEPT/AGENCY: AN10
CASE/REPORT: 08-99999
TEST TYPE: V
ALCO TARGET VALUE: .079
ALCO S/N: X301605

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.50 in

ALCO TARGET	.080	08:49
BLANK TEST	.000	08:51
INTERNAL STANDARD	VERIFIED	08:51
ALCO TV 30.48 in	.086	08:51
BLANK TEST	.000	08:52
SUBJECT SAMPLE CDB 12/18/08	.000	08:54
BLANK TEST	.000	08:55
ALCO TV 30.50 in	.086	08:55
BLANK TEST	.000	08:56

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130288

DECEMBER 14, 2008
TIME 09:05

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	30.50 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~