

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130 289</u>
Name	<u>James Kimura</u>	ID# <u>5677</u> Date <u>2/15/11</u>
A Agency	<u>Alaska State Troopers</u>	Phone # <u>225-5118</u>
Instrument Location	<u>Ketchikan AST Post</u>	
B Alco S/N	<u>78260</u>	Target Value <u>.081</u> High Pressure <u>700 psi</u>
Alco Test Values	<u>.082</u> 1st Alco	<u>.082</u> 2 <sup>nd</sup> Alco
Signature	<u>[Handwritten Signature]</u>	
(OVER)		

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Handwritten Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 27 day of MAY, 2011.

[Handwritten Signature] (Notary Seal Stamp)  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



CRF  
4/21/11

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130289

CRF  
4/21/11

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130289

FEBRUARY 15, 2011

CRF  
4/21/11

OPERATOR'S NAME:

KIMURA, JAMES/K

OPERATOR'S NUMBER: 5677

SUBJECT'S LAST NAME:

VOX

SUBJECT'S FIRST NAME/MI :

VOV

O.L. #: 000000

DEPT/AGENCY: KETA

CASE/REPORT: 11-00000

TEST TYPE: V

ALCO TARGET VALUE: .001

ALCO S/N: 78268

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 29.59 in		
ALCO TARGET	.000	23:13
BLANK TEST	.000	23:14
INTERNAL STANDARD	VERIFIED	23:14
ALCO TV 29.59 in	.002	23:14
BLANK TEST	.000	23:15
SUBJECT SAMPLE	.000	23:15
BLANK TEST	.000	23:16
ALCO TV 29.59 in	.002	23:16
BLANK TEST	.000	23:17

CRF  
4/21/11

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130289

FEBRUARY 15, 2011

TIME 23:18

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	43c
BREATH TUBE:	42c
BAROMETER:	29.59 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcde!ghijklmno  
pqrstuvwxyz{|}~\*