

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130 289</u>
Name	<u>James Kimura</u>	ID# <u>5677</u> Date <u>4/14/2011</u>
A Agency	<u>Alaska State Troopers</u>	Phone # <u>225-5118</u>
Instrument Location	<u>Ketchikan Post</u>	
B Alco S/N	<u>782868</u> JXK	Target Value <u>.081</u> High Pressure <u>600 psi</u>
	<u>78268</u>	
Alco Test Values	<u>.083</u>	<u>.082</u>
	1st Alco	2nd Alco
Signature	<u>[Signature]</u>	

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 31 day of MAY, 2010 2011

[Signature] (Notary Seal Stamp)
Daniel L. Nicholas Kyle Klopfer
Notary Public, State of Alaska
Commission Expires with Office



CRF
4/21/11

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130289

CRF
4/21/11

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130289

APRIL 11, 2011

CRF 4/21/11

OPERATOR'S NAME:

KIMMURA, JAMES/K

OPERATOR'S NUMBER: 5677

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 000000

DEPT/AGENCY: KETO

CASE/REPORT: 11-00000

TEST TYPE: V

ALCO TARGET VALUE: .081

ALCO S/N: 78268

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 29.93 in
ALCO TARGET .081 23:18
BLANK TEST .000 23:19
INTERNAL STANDARD VERIFIED 23:20
ALCO TV 29.93 in .083 23:20
BLANK TEST .000 23:21
SUBJECT SAMPLE .080 23:21
BLANK TEST .000 23:22
ALCO TV 29.93 in .082 23:22
BLANK TEST .000 23:23

CRF
4/21/11

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130289

APRIL 11, 2011

TIME 23:32

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 43c
BREATH TUBE: 43c
BAROMETER: 29.93 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!##%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcde{ghijklmno
pqrstuvwxyz{|}~