

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130289</u>
Name	<u>Byron D. Redburn</u>	ID# <u>2087</u> Date <u>10-28-08</u>
A Agency	<u>Nome PD</u>	Phone # <u>443-5262</u>
Instrument Location <u>Nome PD</u>		
B Alco S/N <u>X301548</u> Target Value <u>.079</u> High Pressure <u>1000</u>		
Alco Test Values		
	<u>.078</u> 1 st Alco	<u>.078</u> 2 nd Alco
Signature <u>[Signature]</u>		<u>208</u> <u>11/4/08</u>

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that:

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Byron D Redburn, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



[Signature]

11/5/08

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

Date

BT9 06/03

NOV 03 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130 289

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130289

OCTOBER 28, 2008 *CS 11/1/08*

OPERATOR'S NAME:

REDBURN BYRON D

OPERATOR'S NUMBER: 2087

SUBJECT'S LAST NAME:

NON DRINKING

SUBJECT'S FIRST NAME/MI :

A

O.L. #: A

DEPT/AGENCY: ONE1

CASE/REPORT: A

TEST TYPE: V

ALCO TARGET VALUE: .079

ALCO S/N: X301548

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.91 in		
ALCO TARGET	.078	09:47
BLANK TEST	.000	09:48
INTERNAL STANDARD	VERIFIED	09:48
ALCO TV 29.91 in	.078 ✓	09:48
BLANK TEST	.000	09:49
SUBJECT SAMPLE	.000 <i>CS</i>	09:50
BLANK TEST	.000 <i>11/1/08</i>	09:51
ALCO TV 29.91 in	.078 ✓	09:51
BLANK TEST	.000	09:52

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130289

OCTOBER 28, 2008

TIME 09:53

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	42c
BAROMETER:	29.91 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHI
HIJKLMNPOQRSTUVWXYZ[\]^_`abcde{ghijklmno
pqrstuvwxyz{|}~"