VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DEC 3 1 2008

DataMaster cdm S/N 130289

Supervisor/Operator Performing the Verification Procedure:

A

Name Byron D. Redburn ID# 2087 Date 12-25-08

Agency Nome PD Phone # 443-5262

Instrument Location Nome PD

B

Alco S/N ~~X301~~ X301548 (1-2-09 BDR) Target Value .079 High Pressure 900

Alco Test Values .077 (1st Alco) .077 (2nd Alco)

Signature B. D. Redburn 1/6/09

(OVER)

(Do Not write in the area below)

nder the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Byron D Redburn , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

1/6/09

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

Date

BT9 06/03

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130289 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130289

DECEMBER 25, 2008 ✓ 1/6/09

OPERATOR'S NAME:
REIDBURN BYRON D
OPERATOR'S NUMBER: 2087
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
A
O.L. #: A
DEPT/AGENCY: OME1
CASE/REPORT: A
TEST TYPE: V
ALCO TARGET VALUE: .079 ✓
ALCO S/N: X301548

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.80 in

ALCO TARGET	.078	15:44
BLANK TEST	.000	15:45
INTERNAL STANDARD	VERIFIED	15:45
ALCO TV 29.80 in	.077 ✓	15:45
BLANK TEST	.000	15:46
SUBJECT SAMPLE	.000	15:47
BLANK TEST	.000	15:48
ALCO TV 29.80 in	.077 ✓	15:48
BLANK TEST	.000	15:49

1/6/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130289

DECEMBER 25, 2008 ✓
TIME 15:41

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	41c
BAROMETER:	29.80 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~