

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

JUL 13 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
Name <u>James Kimura</u>	DataMaster cdm S/N <u>130 289</u> ✓
ID# <u>5677</u> ✓	Date <u>7/9/10</u>
Agency <u>Alaska State Troopers</u>	Phone # <u>(907) 225-5118</u>
Instrument Location <u>Ketchikan AST Post</u>	
Alco S/N <u>78268</u> ✓ Target Value <u>.081</u> ✓ High Pressure <u>1150</u>	
Alco Test Values	
<u>.083</u> ✓ 1 st Alco	<u>.083</u> ✓ 2 nd Alco
Signature <u>[Handwritten Signature]</u>	

(OVER)

BMB
7/15/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Handwritten Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 2th day of July, 2010.

[Handwritten Signature]
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130 289

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130269

JULY 09, 2010

OPERATOR'S NAME:

KIMURA, JAMES/K

OPERATOR'S NUMBER: 5677

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/NI :

VOC

O.L. #: 0000000

DEPT/AGENCY: KET9

CASE/REPORT: 10-000000

TEST TYPE: V

ALCO TARGET VALUE: .081

ALCO S/N: 78268

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 30.19 in
ALCO TARGET .081 22:24
BLANK TEST .000 22:25
INTERNAL STANDARD VERIFIED 22:25
ALCO TV 30.19 in .083 22:25
BLANK TEST .000 22:26
SUBJECT SAMPLE .000 22:27
BLANK TEST .000 22:27
ALCO TV 30.17 in .083 22:28
BLANK TEST .000 22:28

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130269

JULY 09, 2010

TIME 22:15

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 43c
BREATH TUBE: 43c
BAROMETER: 30.20 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrs tuvwxyz{|}~?