

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130 289 ✓

Supervisor/Operator Performing the Verification Procedure:
 Name James K. Kimura ID# 5677 ✓ Date 8/26/10 ✓

A
 Agency Alaska State Troopers Phone # 225-5118
 Instrument Location Ketchikan AST Post

B
 Alco S/N 78268 ✓ Target Value .081 ✓ High Pressure 9000
 Alco Test Values .082 ✓ 1st Alco .081 ✓ 2nd Alco
 Signature James K. Kimura

(OVER)

BMB
8/31/10

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 5th day of Oct, 2010.

Daniel L. Nicholas
(Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130289 *SMB 8/31/10*

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130289

AUGUST 26, 2010

OPERATOR'S NAME:

KIMURA, JAMES/K

OPERATOR'S NUMBER: 5677

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 0000000

DEPT/AGENCY: KET0

CASE/REPORT: 10-000000

TEST TYPE: V

ALCO TARGET VALUE: .001

ALCO S/N: 78268

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 29.94 in
ALCO TARGET .001 00:11
BLANK TEST .000 00:12
INTERNAL STANDARD VERIFIED 00:12
ALCO TV 29.94 in .002 00:13
BLANK TEST .000 00:13
SUBJECT SAMPLE *SMB 8/31/10* .000 00:14
BLANK TEST .000 00:14
ALCO TV 29.94 in .001 00:15
BLANK TEST .000 00:15

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130289

AUGUST 26, 2010

TIME 00:16

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/02/09
HEATERS
SAMPLE CHAMBER: 43c
BREATH TUBE: 43c
BAROMETER: 29.94 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde!ghijklmno
pqrstuvwxyz{|}~`