VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130290

Supervisor/Operator Performing the Verification Procedure:

A

Name DAN STURKIE ID# 3559 Date 01-23-09

Agency ANCHORAGE PD Phone # 786-8900

Instrument Location EAGLE RIVER SUBSTATION

B

Alco S/N X124655 Target Value .083 High Pressure 700

Alco Test Value Average .083 (1st Alco) .082 (2nd Alco)

Signature [signature] 1/27/09

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Daniel Sturkie , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

[signature] 2/3/09

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

BT9 01/00

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130290

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130290

JANUARY 23, 2009

CMB 1/27/09

OPERATOR'S NAME:
STURKIE
OPERATOR'S NUMBER: 3559
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
L
O.L. #: L
DEPT/AGENCY: ANC1
CASE/REPORT: L
TEST TYPE: V
ALCO TARGET VALUE: .083
ALCO S/N: X124655

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 30.09 in

ALCO TARGET	.083	04:41
BLANK TEST	.000	04:41
INTERNAL STANDARD	VERIFIED	04:42
ALCO TV 30.09 in	.083	04:42
BLANK TEST	.000	04:43
SUBJECT SAMPLE	.000	04:44
BLANK TEST	.000	04:45
ALCO TV 30.09 in	.082	04:45
BLANK TEST	.000	04:46

CMB 1/27/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130290

JANUARY 23, 2009
TIME 04:36

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	40c
BAROMETER:	30.09 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~