

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


Supervisor/Operator Performing the Verification Procedure:		Datamaster cdm S/N <u>130290</u>	
Name <u>D.R. Sturkie</u>		ID# <u>3559</u>	Date <u>04-12-08</u>
A	Agency <u>Anchorage Police Dept</u>	Phone # <u>786-8900</u>	
Instrument Location <u>Eagle River Substation</u>			
Alco S/N <u>X301605</u>		Target Value <u>1.080</u>	High Pressure <u>1150</u>
B	Alco Test Value Average <u>.081</u>	<u>.081</u> <u>8/27/08</u>	
	1 st Alco	2 nd Alco	
Signature <u>[Handwritten Signature]</u>		<u>COB</u> <u>7/24/08</u>	
(OVER)			

Under the Alaska Rules of Evidence, I certify that, (Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Daniel Sturkie, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

[Handwritten Signature]

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

4/30/08
Date

4.

a. (1 pt) In what chapter of the *Alaska Breath Alcohol Testing Manual* would you find a short list of Breath Supervisor duties?

b. (1 pt) In what chapter of the *Alaska Breath Alcohol Testing Manual* would you find instructions on how to train and document training of a Designated Operator?

5. (2 pts) What is meant by the "target value" of an Alco?