

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130290</u>
Name	<u>J Swartz</u>	ID#	<u>4314</u> Date <u>4/13/09</u>
A	Agency	<u>SCDL</u>	Phone # <u>269-5592</u>
	Instrument Location	<u>SCDL</u>	
B	Alco S/N	<u>SCOTT08</u>	Target Value <u>.080</u> High Pressure <u>800 psi</u>
	Alco Test Values	<u>.080</u> 1 <sup>st</sup> Alco	<u>.079</u> 2 <sup>nd</sup> Alco
	Signature	<u>[Signature]</u>	<u>CSB</u> <u>5/1/09</u>
(OVER)			

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 19<sup>th</sup> day of May, 2009.  
[Signature] (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130290 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

APRIL 13, 2009 ✓  
OPERATOR'S NAME: SWITZ/J CAS 5/11/09  
OPERATOR'S NUMBER: 4314  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/INIT:  
L  
O.L. & I.  
DEPT/AGENCY: AAAI  
CASE/REPORT:  
TEST TYPE: V  
ALCO TARGET VALUE: 0.00 ✓  
ALCO DIV: SCOTT00

— BREATH ANALYSES —

ALCO ADJUSTED FOR 29.99 in		
ALCO TARGET	.000	22:23
BLANK TEST	.000	22:24
INTERNAL STANDARD	VERIFIED	22:25
ALCO IV 29.99 in	.000 ✓	22:26
BLANK TEST	.000	22:27
SUBJECT SAMPLE	.000	22:28
BLANK TEST	.000	22:29
ALCO IV 29.99 in	.079 ✓	22:30
BLANK TEST	.000	22:31

J 4/13/09

## DIAGNOSTIC CHECK

APRIL 13, 2009 ✓  
TIME 22:21

— DIAGNOSTIC CHECK —

COMPUTER: OKAY  
SERIES: OKAY  
SUPPLY DATE: 01/09/09  
SERIES: OKAY  
SERIES NUMBER: 42c  
SERIES TIME: 42c  
METER: 29.99 in  
METER DETECTOR: OKAY  
METER SPEED: OKAY  
METER: OKAY  
FILTERS: OKAY  
SERIES STANDARD: OKAY  
SERIES: OKAY

PRINTER TEST  
"TEST" (0) +, -, / (0) 23/05/2009 22:21:21  
ALCO ADJUSTED FOR 29.99 in  
ALCO TARGET .000