

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130290</u> ✓
A	Name <u>STEVE DUNN</u>	ID# <u>4106</u> Date <u>4-16-09</u> ✓
	Agency <u>APD</u>	Phone # <u>786-8900</u>
	Instrument Location <u>ANCHORAGE JAIL</u> ✓	
B	Alco S/N <u>X173004</u> ✓	Target Value <u>.081</u> ✓ High Pressure <u>900 PSI</u>
	Alco Test Values <u>.080</u> ✓	<u>.083</u> ✓
	1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
	Signature <u>[Signature]</u>	<u>CSB</u> <u>5/13/09</u>

(OVER)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 19<sup>th</sup> day of May, 2009.

[Signature] (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



APR 17 2009

BT9 06/03

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130290 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

S. DUNN #1361

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130290

APRIL 16, 2009 ✓

COB  
4/18/09

OPERATOR'S NAME:  
DUNN/S  
OPERATOR'S NUMBER: 4106  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/NI :  
V  
O.L. #: V  
DEPT/AGENCY: ANCI  
CASE/REPORT: V  
TEST TYPE: V  
ALCO TARGET VALUE: .081 ✓  
ALCO S/N: X173004 ✓

### --- BREATH ANALYSIS ---

.081 ADJUSTED FOR 29.96 in		
ALCO TARGET	.081	23:40
BLANK TEST	.000	23:41
INTERNAL STANDARD	VERIFIED	23:41
ALCO TV 29.96 in	.080 ✓	23:41
BLANK TEST	.000	23:42
SUBJECT SAMPLE	.080	23:42
BLANK TEST	.000	23:43
ALCO TV 29.96 in	.083 ✓	23:43
BLANK TEST	.000	23:44

## DIAGNOSTIC CHECK

S. DUNN #1361

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130290

APRIL 16, 2009 ✓  
TIME 23:38

### --- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.96 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

### PRINTER TEST

! "\$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~