VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130290

Supervisor/Operator Performing the Verification Procedure:

A

Name Douglas S. Henie ID# 3554 Date 4-29-10

Agency APD Phone # 786-8900

Instrument Location Anchorage Jail

B

Alco S/N 58442 Target Value .079 High Pressure 900

Alco Test Values .078 (1st Alco) .077 (2nd Alco)

Signature D. Henie 1277

(OVER)

NJB 5/3/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 8th day of June, 2010.

(Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

Carolyn M. Noland NOTARY PUBLIC STATE OF ALASKA State of Alaska

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

BT9 06/03

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130290

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm [illegible]

APRIL 29, 2010

BMB 5/3/10

OPERATOR'S NAME:
HENIE DOUGLAS :
OPERATOR'S NUMBER: 1277
SUBJECT'S LAST NAME:
TEST VOC
SUBJECT'S FIRST NAME/MI :
VOC
D.L. #: 111111
DEPT/AGENCY: ANC1
CASE REPORT: 10-0000
TEST TYPE: D
ALCO TARGET VALUE: .079
ALCO S/N: 50442 .

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.68 in

ALCO TARGET	.078	02:07
BLANK TEST	.000	02:08
INTERNAL STANDARD	CERTIFIED	02:08
ALCO TV 29.68 in	.079	02:09
BLANK TEST	.000	02:09
SUBJECT SAMPLE	.000	02:09
BLANK TEST	.000	02:10
ALCO TV 29.68 in	.077	02:10
BLANK TEST	.000	02:11

BMB 5/3/10

D. Henie 1277
4-29-10

DIAGNOSTIC CHECK

D. Henie
1277
4/29/10

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm [illegible]

APRIL 29, 2010
TIME 02:03

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.68 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
[illegible]