

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N ~~130290~~ 130290

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130290

MAY 20, 2009 ✓

OPERATOR'S NAME:  
HEMIE DOUGLAS S  
OPERATOR'S NUMBER: 3554  
SUBJECT'S LAST NAME:  
TEST  
SUBJECT'S FIRST NAME/MI:  
TEST  
O.L. #: 0000000  
DEPT/AGENCY: ANCI  
CASE/REPORT: 000000  
TEST TYPE: T  
ALCO TARGET VALUE: .083 ✓  
ALCO S/N: 78966 ✓

### --- BREATH ANALYSIS ---

.083 ADJUSTED FOR 30.29 in ✓  
ALCO TARGET .084 23:51  
BLANK TEST .000 23:52  
INTERNAL STANDARD VERIFIED 23:52  
ALCO TV 30.29 in .082 ✓ 23:53  
BLANK TEST .000 23:54  
SUBJECT SAMPLE .000 23:54  
BLANK TEST .000 23:55  
ALCO TV 30.30 in .082 ✓ 23:55  
BLANK TEST .000 23:56

DA 1277 5-20-09

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130290

MAY 20, 2009 ✓  
TIME 23:49

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 30.29 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@BCDEFG  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcde fghijklmnop  
qrstuvwxyz{|}~

DA 1277 5-20-09

VERIFICATION OF CALIBRATION REPORT  
of DataMaster cdm Breath Test Instrument

MAY 21 2009

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N ~~78966~~ 130290

Name Douglas S. Henic ID# 3554 Date 5-20-09

A Agency APD Phone # 786-9513

Instrument Location Anchorage Jail

---

B Alco S/N 78966 Target Value .083 High Pressure .1200

Alco Test Value Average .082 .082  
1<sup>st</sup> Alco 2<sup>nd</sup> Alco

Signature Douglas S. Henic

(OVER) BMB  
5/22/09

I, Nita J. Bolz, after being first duly sworn, depose and state as follows. (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 27<sup>th</sup> day of May, 2009.

Carolyn M. Noland (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office

