

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130290 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NO

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130290

JULY 09, 2009

OPERATOR'S NAME:

SITZ, DEREK W

OPERATOR'S NUMBER: 4096

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

G.L. #: 1234567

DEPT/AGENCY: ANCI

CASE REPORT: 00-1234

TEST TYPE: V

ALCO TARGET VALUE: .083

ALCO S/N: 78966

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 30.22 in

ALCO TARGET	.083	00:26
BLANK TEST	.000	00:27
INTERNAL STANDARD	VERIFIED	00:27
ALCO TV 30.22 in	.083	00:27
BLANK TEST	.000	00:28
SUBJECT SAMPLE	.000	00:28
BLANK TEST	.000	00:29
ALCO TV 30.22 in	.083	00:29
BLANK TEST	.000	00:30

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130290

JULY 09, 2009

TIME 00:24

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/00/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 30.22 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!*"#%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJK
LMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstu
vwxyz{|}~

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 13 2009

Supervisor/Operator Performing the Verification Procedure:	
Name <u>Derek Sitz</u>	ID# <u>4096</u> Date <u>7-9-09</u>
Agency <u>Anchorage Police Dept.</u>	Phone # <u>786-8300</u>
Instrument Location <u>Anchorage Jail</u>	
DataMaster cdm S/N <u>130290</u>	
Alco S/N <u>78966</u>	Target Value <u>.083</u> High Pressure <u>250 psi</u>
Alco Test Values	<u>.083</u> 1 st Alco <u>.083</u> 2 nd Alco
Signature <u>Derek Sitz</u>	<u>CSB</u> <u>7/13/09</u>

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 21st day of July, 2009.
Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

