

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		Datamaster cdm S/N <u>130290</u> ✓	
Name <u>D.R. Sturkie</u>		ID# <u>3559</u>	Date <u>9-16-08</u> ✓
A	Agency <u>Anchorage Police</u>	Phone # <u>986-2672/8900</u>	
Instrument Location <u>Eagle River Substation</u> ✓			
Alco S/N <u>58427</u> ✓		Target Value <u>.082</u> ✓	High Pressure <u>500</u>
B	Alco Test Value Average <u>.087</u> ✓		<u>.083</u> ✓
	1 st Alco	2 nd Alco	
Signature <u>D.R. Sturkie</u> ✓		CWS 9/22/08	
(OVER)			

(Do Not write in the area below)


Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Daniel Sturkie, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

10/5/08

Date

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130290 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130290

SEPTEMBER 16, 2008

OPERATOR'S NAME:

STURKIE

OPERATOR'S NUMBER: 3559

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/NI :

L

O.L. #: L

DEPT/AGENCY: ANCI

CASE/REPORT: L

TEST TYPE: V

ALCO TARGET VALUE: .082 ✓

ALCO S/N: 58427 ✓

--- BREATH ANALYSIS ---

.082 ADJUSTED FOR 29.65 in

ALCO TARGET	.081	05:04
BLANK TEST	.000	05:05
INTERNAL STANDARD	VERIFIED	05:05
ALCO TV 29.65 in	.087 ✓	05:05
BLANK TEST	.000	05:06
SUBJECT SAMPLE	.083	05:07
BLANK TEST	.000	05:08
ALCO TV 29.67 in	.083 ✓	05:08
BLANK TEST	.000	05:09

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130290

SEPTEMBER 16, 2008 ✓

TIME 05:00

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 40c

BREATH TUBE: 41c

BAROMETER: 29.65 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~`