

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130290

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130290

NOVEMBER 24, 2009

OPERATOR'S NAME:

NULL

OPERATOR'S NUMBER: 6520

SUBJECT'S LAST NAME:

CALIBRATION

SUBJECT'S FIRST NAME/MI :

VERIFICATION/OF

O.L. #: 012345

DEPT/AGENCY: ANCI

CASE/REPORT: 09-999999

TEST TYPE: V

ALCO TARGET VALUE: .085

ALCO S/N: 70020

--- BREATH ANALYSIS ---

.085 ADJUSTED FOR 29.49 in		
ALCO TARGET	.083	23:53
BLANK TEST	.000	23:54
INTERNAL STANDARD	VERIFIED	23:54
ALCO TV 29.49 in	.085	23:55
BLANK TEST	.000	23:55
SUBJECT SAMPLE	.000	23:56
BLANK TEST	.000	23:57
ALCO TV 29.49 in	.085	23:57
BLANK TEST	.000	23:58

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130290

NOVEMBER 24, 2009
TIME 23:58

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 29.49 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~*

VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm Breath Test Instrument
 State of Alaska
 Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: <u>David Noll</u>		DataMaster cdm S/N <u>130290</u> ✓
Name <u>David Noll</u>		ID# <u>6520</u> Date <u>11-24-09</u> ✓
A	Agency <u>APD</u>	Phone # <u>786-8968</u>
Instrument Location <u>Anchorage Jail - North</u> ✓		
Alco S/N <u>78820</u> ✓		Target Value <u>0.085</u> ✓ High Pressure <u>900</u>
Alco Test Values <u>0.085</u> ✓		<u>0.085</u> ✓
		1 st Alco 2 nd Alco
Signature <u>[Signature]</u>		
		NOV 25 2009
(OVER)		

BMB
11/30/09

(Do Not write in the area below)

- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 18th day of Dec., 2009.

[Signature] (Notary Seal Stamp)
 Carolyn M. Noland
 Notary Public, State of Alaska
 Commission Expires with Office

