

# VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		Datamaster cdm S/N <u>130290</u> ✓	
Name <u>DR Sturkie</u>		ID# <u>3559</u>	Date <u>12-19-08</u> ✓
A	Agency <u>APD</u>	Phone # <u>786-8900</u>	
Instrument Location <u>Eagle River Substation</u> ✓			
Alco S/N <u>57447</u> ✓		Target Value <u>.082</u> ✓	High Pressure <u>500</u>
Alco Test Value Average <u>.084</u> ✓		<u>.084</u> ✓	
1 <sup>st</sup> Alco		2 <sup>nd</sup> Alco	
Signature <u>[Signature]</u> ✓		COB 12/30/08	

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Daniel Sturkie, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed



[Signature]

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

12/31/08

Date

DEC 23 2008

BT9 01/00

# VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N

130290 -

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

TEST

ST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130290

DECEMBER 19, 2008

OPERATOR'S NAME:

STURKIE

OPERATOR'S NUMBER: 3559

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI:

L

O.L. #: L

DEPT/AGENCY: ANCI

CASE/REPORT: L

TEST TYPE: V

ALCO TARGET VALUE: .002

ALCO S/N: 57447

--- BREATH ANALYSIS ---

.002 ADJUSTED FOR 30.06 in  
ALCO TARGET .002 02:32  
BLANK TEST .000 02:33  
INTERNAL STANDARD VERIFIED 02:33  
ALCO TV 30.06 in .004 ✓ 02:34  
BLANK TEST .000 02:34  
SUBJECT SAMPLE .000 02:35  
BLANK TEST .000 02:36  
ALCO TV 30.06 in .004 ✓ 02:36  
BLANK TEST .000 02:37

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130290

DECEMBER 19, 2008

TIME 02:25

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 30.06 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~