


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130291</u> ✓
Name	<u>Sean Fielding</u>	ID#	<u>5189</u> ✓
Date	<u>1/5/09</u> ✓		
A	Agency	<u>Denali National Park</u>	Phone # <u>683-9528</u>
	Instrument Location	<u>Denali National Park HQ</u> ✓	
B	Alco S/N	<u>X 301553</u> ✓	Target Value <u>0.083</u> ✓
			High Pressure <u>600</u>
	Alco Test Value Average	<u>0.085</u> ✓	<u>0.085</u> ✓
		1 st Alco	2 nd Alco
	Signature		<u>003</u> <u>1/27/09</u>

(OVER)


Under the Alaska Rules of Evidence, I certify that:
(Do Not write in the area below)


(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Sean Fielding, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,





Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

2/3/09
Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130291

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130291

JANUARY 05, 2009 ✓ *COS 12/10/09*

OPERATOR'S NAME:
FIELDING/SEAN/R
OPERATOR'S NUMBER: 5189
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
A
O.L. #: 1234567
DEPT/AGENCY: DHLB
CASE/REPORT: 010509
TEST TYPE: U
ALCO TARGET VALUE: .083 ✓
ALCO S/N: X301553

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 27.88 in		
ALCO TARGET	.077	16:12
BLANK TEST	.000	16:13
INTERNAL STANDARD	VERIFIED	16:13
ALCO TU 27.88 in	.085 ✓	16:13
BLANK TEST	.000	16:14
SUBJECT SAMPLE	.000	16:15
BLANK TEST	.000	16:16
ALCO TU 27.88 in	.085 ✓	16:16
BLANK TEST	.000	16:17

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130291

JANUARY 05, 2009
TIME: 16:18

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c

BAROMETER: 27.88 in.

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST
!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKLMNOPQRSTUVWXYZ\]^_`abcdefgijklmnop
qrstuvwxyz{|}~