

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130 291</u>
Name	<u>Jaime L. Smith</u>	ID#	<u>6762</u> Date <u>6/3/10</u>
A Agency	<u>Denali National Park</u>	Phone #	<u>683-9525</u>
Instrument Location		<u>Denali National Park - HQ</u>	
B Alco S/N	<u>X301553</u>	Target Value	<u>.083</u> High Pressure <u>450</u>
Alco Test Value Average	<u>.083</u> ✓	<u>.083</u> ✓	
	1 st Alco	2 nd Alco	
Signature	<u>Jaime L. Smith</u> ✓		
(OVER)			

BMB
6/15/10

I, Nita J. Bolz, after being first duly sworn, depose and state as follows (Do not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 2nd day of July, 2010.

Carolyn M. Noland
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130291

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130291

JUNE 03, 2010

OPERATOR'S NAME:
SMITH/JAIME/L

OPERATOR'S NUMBER: 6762

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

X

O.L. #: X

DEPT/AGENCY: DNL8

CASE/REPORT:

TEST TYPE: U

ALCO TARGET VALUE: .083 ✓

ALCO S/N: X301553

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 27.54 in ✓
ALCO TARGET .076 09:50
BLANK TEST .000 09:51
INTERNAL STANDARD VERIFIED 09:51
ALCO TV 27.54 in .083 09:51
BLANK TEST .000 09:52
SUBJECT SAMPLE .000 09:53
BLANK TEST .000 09:53
ALCO TV 27.56 in .083 ✓ 09:54
BLANK TEST .000 09:54

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130291

JUNE 03, 2010 ✓

TIME 09:46

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 27.56 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!""%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNPOQRSTUVWXYZ\]^_`abcd efghijklmno
pqrstuvwxyz{|}~■