


# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130291 ✓			
Name	Sean Fielding	ID#	5189	Date	10/15/08 ✓	
A	Agency	Denali National Park	Phone #	683-9528		
	Instrument Location	Denali National Park - HQ			✓	
B	Alco S/N	X301686 ✓	Target Value	.086 ✓	High Pressure	900 ✓
	Alco Test Value Average	.084 ✓		.084 ✓		
		1 <sup>st</sup> Alco		2 <sup>nd</sup> Alco		
	Signature					
		(OVER)				

Under the Alaska Rules of Evidence, I certify that:  
(Do Not write in the area below)


(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Sean Fielding, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



  
\_\_\_\_\_  
Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

11/5/08  
\_\_\_\_\_  
Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130291

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130291

OCTOBER 15, 2008 ✓  
*CSB 11/4/08*

OPERATOR'S NAME:  
FIELDING/SEAN/R  
OPERATOR'S NUMBER: 5189  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI:

A  
O.L. #: 1234567  
DEPT/AGENCY: DMLB  
CASE/REPORT: 101508  
TEST TYPE: V  
ALCO TARGET VALUE: .086 ✓  
ALCO S/N: X301686 ✓

--- BREATH ANALYSIS ---

.086 ADJUSTED FOR 27.82 in  
ALCO TARGET .079 08:37  
BLANK TEST .000 08:38  
INTERNAL STANDARD VERIFIED 08:39  
ALCO TU 27.80 in .084 ✓ 08:39  
BLANK TEST .000 08:40  
SUBJECT SAMPLE .000 *CSB* 08:40  
BLANK TEST .000 *CSB* 08:41  
ALCO TU 27.82 in .084 ✓ 08:41  
BLANK TEST .000 08:42

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130291

OCTOBER 15, 2008 ✓  
TIME 08:43

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 40c  
BAROMETER: 27.82 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

! " # % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { } ~ #