

FEB 07 2011 *2011*

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument  
State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130292 ✓
Name	TRACY MURPHY	ID#	7752 ✓
		Date	2-2-11 ✓
A	Agency	JUNEAU POLICE DEPT	Phone #
	Instrument Location	JUNEAU POLICE DEPT	
Alco S/N		X172933 ✓	Target Value
			.078 ✓
			High Pressure
			29.93
Alco Test Values		.081 ✓	.081 ✓
		1st Alco	2nd Alco
Signature		<i>[Signature]</i>	

(OVER)

*BMB  
3/2/11*

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

*[Signature]*  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 16<sup>th</sup> day of MAR, 2011 *2011*

*[Signature]*  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130292

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130292

FEBRUARY 02, 2011

OPERATOR'S NAME: *AMB 3/6/11*  
MURPHY/TRACY/E  
OPERATOR'S NUMBER: 7252  
SUBJECT'S LAST NAME:  
MURPHY  
SUBJECT'S FIRST NAME/MI :  
TRACY/E  
O.L. #: VOC  
DEPT/AGENCY: JNU1  
CASE/REPORT: VOC  
TEST TYPE: V  
ALCO TARGET VALUE: .078  
ALCO S/N: X172933

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.93 in  
ALCO TARGET .078 19:03  
BLANK TEST .000 19:04  
INTERNAL STANDARD VERIFIED 19:04  
ALCO TV 29.93 in .081 19:05  
BLANK TEST .000 19:05  
SUBJECT SAMPLE *AMB 3/6/11* .000 19:06  
BLANK TEST .000 19:06  
ALCO TV 29.94 in .081 19:07  
BLANK TEST .000 19:07

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130292

FEBRUARY 02, 2011  
TIME 19:01

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHAMBER: 48c  
BREATH TUBE: 41c  
BAROMETER: 29.94 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST  
!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGH  
IJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnop  
qrstuvwxyz{|}~\*~