

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130292</u> ✓
Name		<u>Gary Lowe</u>	ID# <u>7219</u> Date <u>10/6/08</u> ✓
A	Agency	<u>Haines P.D.</u>	Phone # <u>766-2121</u>
Instrument Location		<u>Haines P.D.</u> ✓	
Alco S/N		<u>57324</u> ✓	Target Value <u>.080</u> ✓ High Pressure <u>350</u>
B	Alco Test Values	<u>.081</u> ✓ 1 st Alco	<u>.080</u> ✓ 2 nd Alco
Signature		<u>Gary Lowe</u> ✓	<u>CS</u> <u>10/23/08</u>

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Gary A Lowe, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

10/24/08

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130292 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130292

OCTOBER 06, 2008 ✓ *COB 10/23/08*

OPERATOR'S NAME:
LOWE/GARY/A
OPERATOR'S NUMBER: 7219
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC

O.L. #: 123456
DEPT/AGENCY: HNS1
CASE/REPORT: 00-100600
TEST TYPE: U
ALCO TARGET VALUE: .000 ✓
ALCO S/N: 57324 ✓

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.99 in		
ALCO TARGET	.000	15:53
BLANK TEST	.000	15:55
INTERNAL STANDARD	VERIFIED	15:55
ALCO TO 29.99 in	.001 ✓	15:55
BLANK TEST	.000	15:56
SUBJECT SAMPLE	.000 <i>COB</i>	15:56
BLANK TEST	.000 <i>10/23/08</i>	15:57
ALCO TO 29.99 in	.000 ✓	15:57
BLANK TEST	.000	15:58

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130292

OCTOBER 06, 2008 ✓
TIME 15:35

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	29.99 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!""#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJK
LMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~■