

MAR 03 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130293^v

Name Brian Mitchell ID# 4119 Date 2-28-08

A Agency Alaska State Troopers Phone # 486-4121

Instrument Location Kodiak Post, Interview room

B Alco S/N X124630^v Target Value .076^v High Pressure 250

Alco Test Value Average .078 .079

1st Alco 2nd Alco

Signature Brian Mitchell CMB
3/10/08


(OVER)

Under the Alaska Rules of Evidence, I certify that: (Do Not write in the area below)

(a) I am Jeanne Swartz, Criminalist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Brian Mitchell, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

03/10/08
Date

MAR 07 2008

BT9 03/00

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130293

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130293

FEBRUARY 28, 2008

OPERATOR'S NAME: *CAB 3/10/08*
SUPERVISOR
OPERATOR'S NUMBER: 4119
SUBJECT'S LAST NAME:
VERIFICATION
SUBJECT'S FIRST NAME/MI :
VERIFICATION

D.L. #: X
DEPT/AGENCY: KOD0
CASE/REPORT: X
TEST TYPE: U
ALCO TARGET VALUE: .076
ALCO S/N: X124630

--- BREATH ANALYSIS ---

.076 ADJUSTED FOR 29.55 in
ALCO TARGET .075 05:06
BLANK TEST .000 05:07
INTERNAL STANDARD VERIFIED 05:07
ALCO TV 29.55 in .078 05:08
BLANK TEST .000 05:09
SUBJECT SAMPLE .000 *CAB 3/10/08* 05:09
BLANK TEST .000 05:10
ALCO TV 29.55 in .079 05:10
BLANK TEST .000 05:11

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130293

FEBRUARY 28, 2008
TIME 05:12

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/28/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.55 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~