

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130293</u> ✓
Name	<u>Brian Mitchell</u>	ID# <u>4119</u> Date <u>9-7-08</u> ✓
A Agency	<u>Alaska State Troopers</u>	Phone # <u>486-4121</u>
Instrument Location	<u>Kodiak Post, Interview Room</u> ✓	
B Alco S/N	<u>X124626</u> ✓	Target Value <u>.082</u> ✓ High Pressure <u>550</u>
Alco Test Values	<u>.083</u> ✓ 1 <sup>st</sup> Alco	<u>.084</u> ✓ 2 <sup>nd</sup> Alco
Signature	<u><i>Brian Mitchell</i></u> ✓	<u>CSB</u> <u>9/11/08</u>

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Brian Mitchell, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

*Jeanne Swartz*

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

9/11/08

Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130293 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130293

SEPTEMBER 07, 2008 *COB 9/11/08*

OPERATOR'S NAME:  
MITCHELL BRIAN S  
OPERATOR'S NUMBER: 4119  
SUBJECT'S LAST NAME:  
VERIFICATION  
SUBJECT'S FIRST NAME/MI :  
OF CALIBRATION  
O.L. #: XXX  
DEPT/AGENCY: KOD0  
CASE/REPORT: XXX  
TEST TYPE: U  
ALCO TARGET VALUE: .082 ✓  
ALCO S/N: X124626 ✓

--- BREATH ANALYSIS ---

.082 ADJUSTED FOR 29.65 in		
ALCO TARGET	.081	01:20
BLANK TEST	.000	01:21
INTERNAL STANDARD	VERIFIED	01:21
ALCO TV 29.65 in	.083 ✓	01:21
BLANK TEST	.000	01:22
SUBJECT SAMPLE	.000 <i>COB 9/11/08</i>	01:23
BLANK TEST	.000	01:23
ALCO TV 29.65 in	.084 ✓	01:24
BLANK TEST	.000	01:24

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130293

SEPTEMBER 07, 2008 ✓  
TIME 01:25

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	42c
BAROMETER:	29.65 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

! "# \$ % ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~