

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>170293</u> ✓
Name	<u>Brian Mitchell</u>	ID#	<u>4119</u> Date <u>11-22-08</u> ✓
A Agency		<u>Alaska State Troopers</u>	Phone # <u>486-4121</u>
Instrument Location <u>Kodiak Post, Interview Room.</u> ✓			
B Alco S/N		<u>X124626</u> ✓	Target Value <u>.082</u> ✓ High Pressure <u>500</u>
Alco Test Value Average		<u>.083</u> ✓	<u>.084</u> ✓
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature		<u>Brian Mitchell</u> ✓	<u>8/14/08</u>

(OVER)


Under the Alaska Rules of Evidence, I certify that, (Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Brian Mitchell, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Swartz

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

12/2/08

Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130293 ✓

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130293

NOVEMBER 22, 2008 *12/2/08*

OPERATOR'S NAME:  
SUPERVISOR  
OPERATOR'S NUMBER: 4119  
SUBJECT'S LAST NAME:  
MITCHELL B S  
SUBJECT'S FIRST NAME/MI :  
XXX *VOC - BSMC*  
O.L. #: XXX  
DEPT/AGENCY: K000  
CASE/REPORT: XXX  
TEST TYPE: U  
ALCO TARGET VALUE: .082 ✓  
ALCO S/N: X124626

### --- BREATH ANALYSIS ---

.082 ADJUSTED FOR 29.52 in		
ALCO TARGET	.080	23:05
BLANK TEST	.000	23:06
INTERNAL STANDARD	VERIFIED	23:06
ALCO TV 29.52 in	.083 ✓	23:06
BLANK TEST	.000	23:07
SUBJECT SAMPLE <i>12/2/08</i>	.080	23:07
BLANK TEST	.000	23:08
ALCO TV 29.52 in	.084 ✓	23:08
BLANK TEST	.000	23:09

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130293

NOVEMBER 22, 2008 ✓  
TIME 23:10

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 40c  
BREATH TUBE: 41c  
BAROMETER: 29.52 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF6  
HIJKLMNPOQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqrstuvwxy z{|}~||