

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument*

*State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

DataMaster cdm S/N 130293 ✓

Supervisor/Operator Performing the Verification Procedure:

Name Colleen O'Bryant ID# 6237 Date 12/7/10 ✓

A Agency SCDL Phone # 269-5440

Instrument Location SCDL

---

B Alco S/N SC0408 ✓ Target Value 0.080 High Pressure 200psi

Alco Test Values 0.078 ✓ 0.079 ✓  
1<sup>st</sup> Alco 2<sup>nd</sup> Alco

Signature C. O'Bryant

(OVER)

*BMB  
12/16/10*

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

*Nita J. Bolz*  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 18<sup>th</sup> day of JAN, 2010 *DLN*

*[Signature]*  
 (Notary Seal Stamp)  
 Daniel L. Nicholas  
 Notary Public, State of Alaska  
 Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130293

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130293

DECEMBER 07, 2010

OPERATOR'S NAME:  
O'BRYEN/COLLEEN/S  
OPERATOR'S NUMBER: 6237  
SUBJECT'S LAST NAME:  
UO

SUBJECT'S FIRST NAME/MI :  
UOC

O.L. #: 0123456  
DEPT/AGENCY: AAA1  
CASE/REPORT: 10-1234  
TEST TYPE: U

ALCO TARGET VALUE: .080  
ALCO S/N: 970730

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 29.93 in  
ALCO TARGET .080 12:39  
BLANK TEST .000 12:40  
INTERNAL STANDARD VERIFIED 12:40  
ALCO TV 29.93 in .078 12:40  
BLANK TEST .000 12:41  
SUBJECT SAMPLE .080 12:42  
BLANK TEST .000 12:43  
ALCO TV 29.91 in .079 12:43  
BLANK TEST .000 12:44

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130293

DECEMBER 07, 2010  
TIME 12:29

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 43c  
BAROMETER: 29.94 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~