

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

MAR 16 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130294</u> ✓	
Name <u>J Swartz</u>		ID# <u>4314</u>	Date <u>2/28/09</u> ✓
A	Agency <u>SCDL</u>	Phone # <u>269-5592</u>	
Instrument Location <u>SCDL</u>			
Alco S/N <u>SC07108</u> ✓		Target Value <u>.080</u> ✓	High Pressure <u>900 psi</u>
B	Alco Test Values	<u>.074</u> ✓ 1 st Alco	<u>.077</u> ✓ 2 nd Alco
Signature <u>J Swartz</u>		COB <u>3/24/09</u>	
(OVER)			

Under the Alaska Rules of Evidence, I certify that (copy of value in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Jeanne M Swartz, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

J Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

3/24/09

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster edm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

130294 ✓
DataMaster edm S/N 13040 *J*

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER edm 130294

FEBRUARY 28, 2009 ✓

OPERATOR'S NAME: SMARTZ/J *COB 3/2/09*

OPERATOR'S NUMBER: 4314

SUBJECT'S LAST NAME: VERIFICATION OF CAL

SUBJECT'S FIRST NAME/NI: L

D.L. #: L

DEPT/AGENCY: AAAI

CASE/REPORT: TEST TYPE: U

ALCO TARGET VALUE: .080 ✓

ALCO S/N: SCOTT08 ✓

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 30.16 in		
ALCO TARGET	.080	09:12
BLANK TEST	.000	09:13
INTERNAL STANDARD	VERIFIED	09:13
ALCO TV 30.16 in	.074 ✓	09:13
BLANK TEST	.000	09:14
SUBJECT SAMPLE <i>COB 3/2/09</i>	.000	09:14
BLANK TEST	.000	09:15
ALCO TV 30.16 in	.077 ✓	09:15
BLANK TEST	.000	09:16

J 2/25/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER edm 130294

FEBRUARY 28, 2009
TIME 09:09

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 43c

BREATH TUBE: 42c

BAROMETER: 30.16 in

FLOW DETECTOR: OKAY

RUMP HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=?@ABCDEF
HI JKLMNOPQRSTU VWXYZ[\]^_`'abcde fghijklmno
pqrstuvwxyz{|}~