

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument  
State of Alaska

MAR 04 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130294</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>C. RITALA</u>	ID# <u>3791</u> Date <u>3/1/09</u>
Agency <u>APD</u>	Phone # <u>766-2640</u>
Instrument Location <u>Wollam Sub (Eagle River)</u>	
Alco S/N <u>X124655</u> Target Value <u>.083</u> High Pressure <u>500</u>	
Alco Test Value Average <u>.083</u> <u>.083</u>	
1 <sup>st</sup> Alco 2 <sup>nd</sup> Alco	
Signature <u>[Handwritten Signature]</u> <u>X349</u> <u>OB</u> <u>3/24/09</u>	
(OVER)	

Under the Alaska Rules of Evidence, I certify that (write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Chris Ritala, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



[Handwritten Signature]

3/24/09

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130294

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130294

MARCH 01, 2009

OPERATOR'S NAME:

RITALA

OPERATOR'S NUMBER: 3791

SUBJECT'S LAST NAME:

VERIFICATION OF CALIBRATION

SUBJECT'S FIRST NAME/MI :

A

OIL #: A

DEPT/AGENCY: ANCI

CASE/REPORT: A

TEST TYPE: U

ALCO TARGET VALUE: 083

ALCO S/N: X124655

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 29.85 in		
ALCO TARGET	.082	00:58
BLANK TEST	.000	00:59
INTERNAL STANDARD	VERIFIED	00:59
ALCO TV 29.85 in	.083	00:59
BLANK TEST	.000	01:00
SUBJECT SAMPLE	.000	01:00
BLANK TEST	.000	01:01
ALCO TV 29.86 in	.083	01:01
BLANK TEST	.000	01:02

CR/1349  
3/1/09

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130294

MARCH 01, 2009

TIME 00:56

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 42c

BAROMETER: 29.85 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! "#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ\]^\_`abcde fghijklmno  
pqrstuvwxyz{|}~■

CR/1349  
3/1/09