

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

APR 28 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130294</u>
Name <u>Jean Achee</u>		ID# <u>1470</u> Date <u>April 23, 2010</u>
A	Agency <u>Wasilla Police Department</u>	Phone # <u>352-5401</u>
	Instrument Location <u>Wasilla Police Department</u>	
Alco S/N <u>G30491</u>		Target Value <u>0.83</u> High Pressure <u>750</u>
B	Alco Test Value Average	<u>.082</u> <u>.081</u>
	1 st Alco	2 nd Alco
Signature	<u>Jean Achee</u>	
(OVER)		

BMB
5/3/10

I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 18th day of June, 2010.

Carolyn M. Noland (Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130294

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130294

APRIL 23, 2010

OPERATOR'S NAME: *JP* *BMB*
ACHEE, JEAN P
OPERATOR'S NUMBER: 1470
SUBJECT'S LAST NAME: *5/3/10*

UOC
SUBJECT'S FIRST NAME/MI :
UOC

O.L. #: 1234567890
DEPT/AGENCY: WASI
CASE/REPORT: 100423
TEST TYPE: V
ALCO TARGET VALUE: .083
ALCO S/N: 630491

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 30.03 in		
ALCO TARGET	.083	10:06
BLANK TEST	.000	10:07
INTERNAL STANDARD	VERIFIED	10:07
ALCO TU 30.03 in	.082	10:07
BLANK TEST	.000	10:08
SUBJECT SAMPLE	.000	10:09
BLANK TEST	.000	10:09
ALCO TU 30.03 in	.081	10:10
BLANK TEST	.000	10:10

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130294

APRIL 23, 2010
TIME 10:04

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	30.03 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!""&?()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcde fghi jklmno
pqrstuvwxyz{|}~■