

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

JUL 23 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130294
A	Name	C. KIRALA	ID# 3791 Date 7/21/09 ✓
	Agency	APD	Phone # 786-2640
	Instrument Location	New Wollam Job Station (Eagle River)	✓
B	Alco S/N	X172946 ✓	Target Value 084 ✓ High Pressure 275
	Alco Test Value Average	081 ✓	081 ✓
		1 st Alco	2 nd Alco
	Signature	Chris Rind / 1349 ✓	COB 8/6/09


(OVER)


(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 14th day of Aug, 2009.
 (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130294 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130294

JULY 21, 2009 ✓

OPERATOR'S NAME: *CRB 8/16/09*
RITALA

OPERATOR'S NUMBER: 3791

SUBJECT'S LAST NAME:

VERIFICATION OF CALIBRATION

SUBJECT'S FIRST NAME/NI :

MOVED TO NEW WOLLAM SUB IN ER

O.L. #: A

DEPT/AGENCY: ANCI

CASE/REPORT: A

TEST TYPE: U

ALCO TARGET VALUE: .004

ALCO S/N: X172946 ✓

--- BREATH ANALYSIS ---

.004 ADJUSTED FOR 29.41 in
ALCO TARGET .002 22:31
BLANK TEST .000 22:31
INTERNAL STANDARD VERIFIED 22:32
ALCO TV 29.41 in .001 ✓ 22:32
BLANK TEST .000 22:33
SUBJECT SAMPLE *CRB 8/16/09* .000 22:33
BLANK TEST .000 22:34
ALCO TV 29.41 in .001 ✓ 22:34
BLANK TEST .000 22:35

CRB 8/16/09 7/21/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130294

JULY 21, 2009 ✓
TIME 22:28

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 39c
BAROMETER: 29.41 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~

CRB 8/16/09 7/21/09