

SEP 10 2009

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument  
State of Alaska*

**Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program**

Supervisor/Operator Performing the Verification Procedure: <span style="float: right;">DataMaster cdm S/N <u>130294</u> ✓</span>	
Name <u>Brandi Barnett</u>	ID# <u>7038</u> Date <u>8/6/09</u>
A Agency <u>SCDL</u> Phone # <u>269-5695</u>	
Instrument Location <u>SCDL</u>	
B Alco S/N <u>SCOTT08</u> Target Value <u>0.080</u> High Pressure <u>650 PSI</u>	
Alco Test Values <u>0.077</u> <u>0.077</u>	
1 <sup>st</sup> Alco 2 <sup>nd</sup> Alco	
Signature <u>Brandi M. Barnett</u> <span style="float: right;">CBS 10/27/09</span>	
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 9<sup>th</sup> day of NOV, 2009.

Carolyn M. Noland  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130294 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130294

AUGUST 06, 2009 ✓

OPERATOR'S NAME: *CDL 10/2/09*  
BARNETT/B/M  
OPERATOR'S NUMBER: 7038  
SUBJECT'S LAST NAME:  
UOC  
SUBJECT'S FIRST NAME/MI :  
UOC  
O.L. #: UOC  
DEPT/AGENCY: AAA1  
CASE/REPORT: UOC  
TEST TYPE: U  
ALCO TARGET VALUE: 0.080 ✓  
ALCO S/N: SCOTT08

### --- BREATH ANALYSIS ---

.080 ADJUSTED FOR 29.90 in  
ALCO TARGET .079 14:17  
BLANK TEST .000 14:18  
INTERNAL STANDARD VERIFIED 14:18  
ALCO TV 29.90 in .077 14:18  
BLANK TEST .000 14:19  
SUBJECT SAMPLE *CDL 10/2/09* .000 14:20  
BLANK TEST .000 14:20  
ALCO TV 29.91 in .077 14:21  
BLANK TEST .000 14:21

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130294 ✓

AUGUST 06, 2009 ✓  
TIME 14:07

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHAMBER: 42c  
BREATH TUBE: 26c  
BAROMETER: 29.88 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
qrstuvwxyz{|}~■