

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130295 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130295

APRIL 15, 2009 ✓ *COB 5/13/09*

OPERATOR'S NAME: MARTIN, PATRICK
OPERATOR'S NUMBER: 4072
SUBJECT'S LAST NAME: MARTIN
SUBJECT'S FIRST NAME/NI: PATRICK
O.L. #: 1234567
DEPT/AGENCY: ANCI
CASE/REPORT: UOC
TEST TYPE: U
ALCO TARGET VALUE: .000 ✓
ALCO S/N: 31085 ✓

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.12 in		
ALCO TARGET	.000	15:32
BLANK TEST	.000	15:32
INTERNAL STANDARD	VERIFIED	15:33
ALCO TV 30.12 in	.003 ✓	15:33
BLANK TEST	.000	15:34
SUBJECT SAMPLE <i>COB 5/13/09</i>	.000	15:34
BLANK TEST	.000	15:35
ALCO TV 30.12 in	.004 ✓	15:35
BLANK TEST	.000	15:36

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130295

APRIL 15, 2009 ✓
TIME 15:29

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/09/09
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	42c
BAROMETER:	30.12 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~■

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		Datamaster cdm SN <u>130295</u> ✓	
Name <u>Patrick E. Martin</u>		ID# <u>4072</u>	Date <u>4-15-09</u> ✓
A	Agency <u>APD</u>	Phone # <u>796-2601</u>	
Instrument Location <u>Seely Sub Station</u>			
Alco S/N <u>31085</u> ✓		Target Value <u>.080</u> ✓	High Pressure <u>550 psi</u>
B	Alco Test Value Average <u>.083</u> ✓	<u>.084</u> ✓	
	1 st Alco	2 nd Alco	
Signature <u>Patrick E. Martin</u> ✓			<u>CRB</u> <u>5/13/09</u>
(OVER)			

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 19th day of May 2009.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



BT9 01/00

APR 17 2009