

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130295<sup>✓</sup>

Name Joe Corbett ID# 4971<sup>✓</sup> Date 8/30/10<sup>✓</sup>

A Agency Bethel PD Phone # 545-0139

Instrument Location Bethel PD

---

B Alco S/N X301563<sup>✓</sup> Target Value .079<sup>✓</sup> High Pressure 400 psi

Alco Test Values .081<sup>✓</sup> .081<sup>✓</sup>

1<sup>st</sup> Alco                      2<sup>nd</sup> Alco

Signature [Handwritten Signature]

(OVER)

*[Handwritten initials]*  
9/3/10

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Handwritten Signature]  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 9<sup>th</sup> day of SEPT, 2010.

[Handwritten Signature] (Notary Seal Stamp)  
 Daniel L. Nicholas  
 Notary Public, State of Alaska  
 Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130295

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130295

AUGUST 30, 2010

OPERATOR'S NAME:

CORBETT/JOE

OPERATOR'S NUMBER: 4971

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

UOC

O.L. #: 1234567

DEPT/AGENCY: BETI

CASE/REPORT: UOC 8/30/10

TEST TYPE: T

ALCO TARGET VALUE: .079

ALCO S/N: X301563

— BREATH ANALYSIS —

.079 ADJUSTED FOR 30.09 in  
ALCO TARGET .079 19:07  
BLANK TEST .000 19:08  
INTERNAL STANDARD VERIFIED 19:08  
ALCO TV 30.09 in .001 19:09  
BLANK TEST .000 19:10  
SUBJECT SAMPLE .000 19:10  
BLANK TEST .000 19:11  
ALCO TV 30.09 in .001 19:11  
BLANK TEST .000 19:12

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130295

AUGUST 30, 2010

TIME 19:05

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 40c

BAROMETER: 30.09 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@BCDEFG  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
qrstuvwxyz{|}~|