VERIFICATION OF CALIBRATION REPORT

State of Alaska
Scientific Crime Detection Laboratory
Statewide Breath Alcohol Program

Intoximeter S/N 130295

Supervisor/Operator Performing the Verification Procedure:

A

Name Joan Howell ID# 7108 Date 09-30-08

Agency Hoonah Police Dept Phone # 907-945-3655

Instrument Location 379 Raven Drive

B

Alco S/N X124669 Target Value .083 High Pressure 500

MASTER PROM _____ Channel 0 _____

MASTER CKSUM _____ Channel 1 _____

Alco Test Value .081 / .080 Channel 2 _____

Channel 3 _____

Signature Joan Howell

10/23/08

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Joan Howell , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

10/24/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

BT6 9/95

please email
verification on Calibration to:
joni_howell@hotmail.com

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130295

SEPTEMBER 30, 2008 CAB 10/23/08

OPERATOR'S NAME:
HOWELL
OPERATOR'S NUMBER: 7108
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
J
O.L. #: K
DEPT/AGENCY: HHH1
CASE/REPORT: J
TEST TYPE: V
ALCO TARGET VALUE: .083
ALCO S/N: X124669

— BREATH ANALYSIS —

.083 ADJUSTED FOR 29.90 in

ALCO TARGET	.082	13:24
BLANK TEST	.000	13:25
INTERNAL STANDARD	VERIFIED	13:25
ALCO TV 29.90 in	.081	13:25
BLANK TEST	.000	13:26
SUBJECT SAMPLE	.000 CAB	13:28
BLANK TEST	.000 10/23/08	13:28
ALCO TV 29.90 in	.080	13:29
BLANK TEST	.000	13:30

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130295

SEPTEMBER 30, 2008
TIME 13:11

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	41c
BAROMETER:	29.90 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~

10/23/08