

VERIFICATION OF CALIBRATION REPORT

Of Data-Master cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program

dps.scdl.toxicology@alaska.gov

APR 26 2011

Data-Master cdm S/N: 130296

Supervisor/Operator Performing the Verification Procedure:

Name John J. Waldron ID: # 3392 Date: 04/25/11

A.

Agency Unalaska Department of Public Safety Phone 907-581-1233

Instrument Location Unalaska Department of Public Safety, PO Box 370, 29 Safety Way, Unalaska, AK 99685

Alco S/N: 77198 Target Value: 0.079 ^{BMB} 5/13/11 ~~-.078~~ High Pressure: 1100

B.

Alco Test Values .080 .081
1st Test Value 2nd Test Value

Signature [Signature]

(OVER)

BMB
5/13/11

(Do not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 31 day of MAY, 2011.

[Signature] (Notary Seal Stamp)

~~Daniel L. Nicholas~~ Kyle Klopfer
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

Of Data-Master cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program

Data-Master cdm S/N: 130296

Supervisor/Operator & Number Performing the Verification Procedure: John J. Waldron, #3392
Department and Date: Unalaska DPS

(CONTINUED FROM FRONT PAGE)

C. TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130296

APRIL 25, 2011

OPERATOR'S NAME:

WALDRON/JOHN/J

OPERATOR'S NUMBER: 3392

SUBJECT'S LAST NAME:

UIC

SUBJECT'S FIRST NAME/MI :

UIC

O.L. #: 0123456789

DEPT/AGENCY: UNK1

CASE/REPORT: 11-00000

TEST TYPE: U

ALCO TARGET VALUE: .079

ALCO S/N: 77198

— BREATH ANALYSIS —

.079 ADJUSTED FOR 29.77 in
ALCO TARGET .078 04:40
BLANK TEST .000 04:41
INTERNAL STANDARD VERIFIED 04:41
ALCO TV 29.75 in .080 04:41
BLANK TEST .000 04:42
SUBJECT SAMPLE .080 04:42
BLANK TEST .000 04:43
ALCO TV 29.77 in .081 04:43
BLANK TEST .000 04:44

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130296

APRIL 25, 2011

TIME 04:46

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 40c
BREATH TUBE: 40c
BAROMETER: 29.77 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLNMOPQRSTUVWXYZ\]^_`abcde fghi.jklmno
pqrstuvwxyz{|}~||