Print Form

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130296

Supervisor/Operator Performing the Verification Procedure:

A

Name Jennifer L. Shockley ID# 4552 Date May 19, 2008

Agency Unalaska Dept. of Public Safety Phone # 907-581-1233

Instrument Location Unalaska Dept. of Public Safety

B

Alco S/N X301573 Target Value .075 High Pressure 300

Alco Test Values .076 (1st Alco) .076 (2nd Alco)

Signature [signature] CJB 6/3/08

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Jennifer L Shockley , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

[signature] 6/5/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

MAY 27 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130296

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130296

MAY 19, 2008 CMB 6/3/08

OPERATOR'S NAME:
SHOCKLEY/JENNIFER/L
OPERATOR'S NUMBER: 4552
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: VOC
DEPT/AGENCY: UNK1
CASE/REPORT: VOC
TEST TYPE: V
ALCO TARGET VALUE: .075
ALCO S/N: X301573

--- BREATH ANALYSIS ---

.075 ADJUSTED FOR 29.90 in

ALCO TARGET	.074	17:34
BLANK TEST	.000	17:35
INTERNAL STANDARD	VERIFIED	17:35
ALCO TV 29.90 in	.076	17:36
BLANK TEST	.000	17:36
SUBJECT SAMPLE	.000 CMB 6/3/08	17:37
BLANK TEST	.000	17:38
ALCO TV 29.90 in	.076	17:39
BLANK TEST	.000	17:39

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130296

MAY 19, 2008
TIME 17:32

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/28/01
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	41c
BAROMETER:	29.90 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgnijklmno
pqrstuvwxyz{|}~