

VERIFICATION OF CALIBRATION REPORT

Of Data-Master cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory -- Statewide Breath Alcohol Program

NOV 6 9 2009

Supervisor/Operator Performing the Verification Procedure:			Data-Master cdm S/N: <u>130296</u>
Name <u>John J. Waldron</u>			ID: # <u>3392</u> Date: <u>11/04/09</u>
A.	Agency <u>Unalaska Department of Public Safety</u>		Phone <u>907-581-1233</u>
Instrument Location <u>Unalaska Department of Public Safety, PO Box 370, 29 Safety Way, Unalaska, AK 99685</u>			
Alco S/N: <u>78966</u>			Target Value: <u>.082</u> High Pressure: <u>1050</u>
B.	Alco Test Values	<u>.081</u>	<u>.081</u>
		1 st Test Value	2 nd Test Value
Signature <u>[Signature]</u>			

(OVER)

BMB
11/4/09


(Do not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 21st day of Dec, 2009.
[Signature]
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT
Of Data-Master cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory -- Statewide Breath Alcohol Program

Data-Master cdm S/N: 130296 ✓

Supervisor/Operator & Number Performing the Verification Procedure: John J. Waldron, #3392 ✓
 Department and Date: Unalaska DPS,

(CONTINUED FROM FRONT PAGE)

C. TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
 DATAMASTER cdm 130296 ✓

NOVEMBER 04, 2009

OPERATOR'S NAME: *BMB*
 WALDRON/JOHN/J
 OPERATOR'S NUMBER: 3392
 SUBJECT'S LAST NAME: *11/14/09*
 UOC
 SUBJECT'S FIRST NAME/MI :
 UOC

O.L. #: 0123456789
 DEPT/AGENCY: UNK1
 CASE/REPORT: 09-00000
 TEST TYPE: U
 ALCO TARGET VALUE: .082 ✓
 ALCO S/N: 78966 ✓

--- BREATH ANALYSIS ---

.082 ADJUSTED FOR 28.97 in		
ALCO TARGET	.079	02:41
BLANK TEST	.000	02:43
INTERNAL STANDARD	VERIFIED	02:43
ALCO TV 28.97 in	.081	02:43
BLANK TEST	.000	02:44
SUBJECT SAMPLE	.000	02:44
BLANK TEST	.000	02:45
ALCO TV 28.97 in	.081	02:46
BLANK TEST	.000	02:46

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
 DATAMASTER cdm 130296 ✓

NOVEMBER 04, 2009 ✓
 TIME 02:47

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
 PROGRAM: OKAY
 SOFTWARE DATE: 02/20/01

HEATERS
 SAMPLE CHAMBER: 49c
 BREATH TUBE: 41c

BAROMETER: 28.97 in

FLOW DETECTOR: OKAY

PUMP
 HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
 MNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
 qrstuvwxyz{|}~■