

# VERIFICATION OF CALIBRATION REPORT

Of Data-Master cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory -- Statewide Breath Alcohol Program

MAR 10 2010

Supervisor/Operator Performing the Verification Procedure: Data-Master cdm S/N: 130296

Name John J. Waldron ID: # 3392 Date: 03/06/10

A. Agency Unalaska Department of Public Safety Phone 907-581-1233

Instrument Location Unalaska Department of Public Safety, PO Box 370, 29 Safety Way, Unalaska, AK 99685

Alco S/N: 78966 Target Value: .082 High Pressure: 900

B. Alco Test Values .084 .085  
1<sup>st</sup> Test Value 2<sup>nd</sup> Test Value

Signature [Handwritten Signature]

(OVER)

*BMB 3/18/2010*

(Do not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 2<sup>nd</sup> day of April, 2010.

Carolyn M. Noland (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

*Of Data-Master cdm Breath Test Instrument  
State of Alaska  
Scientific Crime Detection Laboratory -- Statewide Breath Alcohol Program*

Data-Master cdm S/N: 130296 ✓

Supervisor/Operator & Number Performing the Verification Procedure: John J. Waldron, #3392  
Department and Date: Unalaska DPS,

(CONTINUED FROM FRONT PAGE)

**C. TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.**

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130296  
MARCH 06, 2010

*BMB  
3/18/2010*

OPERATOR'S NAME: WALDRON/JOHN/J  
OPERATOR'S NUMBER: 3392  
SUBJECT'S LAST NAME: UOC  
SUBJECT'S FIRST NAME/MI: UOC  
O.L. #: 0123456789  
DEPT/AGENCY: UNK1  
CASE/REPORT: 10-00000  
TEST TYPE: U  
ALCO TARGET VALUE: .082 ✓  
ALCO S/N: 78966 ✓

--- BREATH ANALYSIS ---

ALCO TARGET	.081	19:49
BLANK TEST	.000	19:50
INTERNAL STANDARD	VERIFIED	19:50
ALCO TU 29.78 in	.084	19:51
BLANK TEST	.000	19:52
SUBJECT SAMPLE	.000	19:52
BLANK TEST	.000	19:53
ALCO TU 29.78 in	.085	19:53
BLANK TEST	.000	19:54

*BMB  
3/18/2010*

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130296  
MARCH 06, 2010 ✓  
TIME 19:56

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 48c  
BREATH TUBE: 41c  
BAROMETER: 29.80 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { } ~ |