

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

MAY 25 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130296 ✓			
Name	Jennifer L. Shockley	ID#	4552	Date	05/21/2010 ✓	
A	Agency	Unalaska Dept. of Public Safety	Phone #	(907) 581-1233		
	Instrument Location	UDPS Interview Room ✓				
Alco S/N		78966 ✓	Target Value	.082 ✓	High Pressure	780
B	Alco Test Values	.085 ✓	.085 ✓			
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco			
Signature	J.L. Shockley ✓		CJB 5/26/10			

(OVER)


(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 24<sup>th</sup> day of June, 2010.

  
(Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130296 ✓

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130296 ✓

MAY 21, 2010 ✓ *ops 5/20/10*

OPERATOR'S NAME: SHOCKLEY/JENNIFER/L  
OPERATOR'S NUMBER: 4552  
SUBJECT'S LAST NAME: VUC  
SUBJECT'S FIRST NAME/MI: VUC  
O.L. #: VUC  
DEPT/AGENCY: UNK1  
CASE/REPORT: VUC  
TEST TYPE: V  
ALCO TARGET VALUE: .002 ✓  
ALCO S/N: 78966 ✓

### --- BREATH ANALYSIS ---

.002 ADJUSTED FOR 30.03 in ✓  
ALCO TARGET .002 21:01  
BLANK TEST .000 21:01  
INTERNAL STANDARD VERIFIED 21:01  
ALCO TU 30.03 in .005 ✓ 21:02  
BLANK TEST .000 21:03  
SUBJECT SAMPLE *ops 5/20/10* .000 21:03  
BLANK TEST .000 21:04  
ALCO TU 30.03 in .005 ✓ 21:04  
BLANK TEST .000 21:05

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130296 ✓

MAY 21, 2010 ✓  
TIME 20:58

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 40c  
BREATH TUBE: 41c  
BAROMETER: 30.01 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

! " # % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { } ~ |