

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130297</u> ✓	
Name		<u>JAMES HOELSCHER</u>	ID# <u>5343</u> Date <u>5-13-08</u> ✓	
A	Agency	<u>Hooper Bay Police Dept.</u>	Phone # <u>758-4665</u>	
Instrument Location		<u>Hooper Bay, AK</u> ✓		
Alco S/N		<u>65859</u> ✓	Target Value <u>.082</u> ✓ High Pressure <u>500 PSI</u>	
B	Alco Test Value Average		<u>.088</u> ✓	<u>.080</u> ✓
		1 st Alco	2 nd Alco	
Signature		<u>[Signature]</u> ✓		<u>CSB</u> <u>11/21/08</u>
(OVER)				

Under the Alaska Rules of Evidence, I certify that: (Do Not Write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, James Hoelscher, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

[Signature]

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

11/21/08
Date

NOV 18 2008

BT9 03700

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130297 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

Breath Test Record

Serial Number 130297 Model DataMaster
Location HPB1 Hooper Bay Police Dept.

COB
11/20/08

Subject Information

Last Name Verification O.L. # N/A
First Name Of Calibration

Operator Information

Operator Name Hoelscher/James/C Case N/A
Operator # 5343 Department HPB1

Test Results

Alco Target .082 Alco Serial 65859
Nominal 0.08 Barometer 29.46 Test Type V
Dry Gas 1 0.088 Time: 14:20:31 Test Date 05/13/2008
Subject Alcohol 0.000 Time: 14:21:58 Test Time 14:18:31
Dry Gas 2 0.080 Time: 14:23:05 Report Y

Diagnostic Test Record

Serial Number 130297 Model DataMaster
Location HPB1 Hooper Bay Police Dept.

Diagnostic Test Results

Test Date 05/13/2008 Computer Okay
Test Time 14:15:58 Program Okay
Software Date 02/20/2001
Sample Chamber (C) 48
Breath Tube (C) 41
Barometer (in) 29.46
Flow Detector Okay
High Speed Pump Okay
Detector Okay
Filter Okay
Quartz Okay
Calibration Okay