

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAY 17 2010

Supervisor/Operator Performing the Verification Procedure: **DataMaster cdm S/N** 130297

Name JAMES HUELSCHER ID# 5343 Date 5-10-10

A Agency HOOPER BAY POLICE DEPT. Phone # 758-4615

Instrument Location Hooper Bay Police Dept.

B Alco S/N X301526 Target Value .080 High Pressure 200 PSI

Alco Test Value Average .079 .078
1st Alco 2nd Alco

Signature [Signature] [Signature]

(OVER)


BMB
5/24/10

- (Do Not write in the area below)
- 1. Nita J. Bolz, after being first duly sworn, depose and state as follows:
 - (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 17th day of June, 2010.

[Signature]
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130297

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130297

MAY 10, 2010

OPERATOR'S NAME: HOELSCHER/JAMES/C
OPERATOR'S NUMBER: 5343
SUBJECT'S LAST NAME: VERIFICATION
SUBJECT'S FIRST NAME/MI :
OF CALIBRATION
O.L. #: N/A
DEPT/AGENCY: HPB1
CASE/REPORT: N/A
TEST TYPE: V
ALCO TARGET VALUE: .000
ALCO S/N: X301526

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.94 in
ALCO TARGET .000 11:27
BLANK TEST .000 11:28
INTERNAL STANDARD VERIFIED 11:29
ALCO TV 29.94 in .079 11:29
BLANK TEST .000 11:30
SUBJECT SAMPLE .000 11:30
BLANK TEST .000 11:31
ALCO TV 29.94 in .078 11:31
BLANK TEST .000 11:32

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130297

MAY 10, 2010
TIME 11:25

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 41c
BAROMETER: 29.94 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~■