AUG 1 3 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130297

Supervisor/Operator Performing the Verification Procedure:

A

Name JAMES HOELSCHER ID# 5343 Date 8/3/10 DLN 8/13/10

Agency Hooper Bay Public Safety Phone # 758-4615

Instrument Location Hooper Bay

B

Alco S/N X172968 Target Value .081 High Pressure 1000 PSI

Alco Test Values .079 (1st Alco) .079 (2nd Alco)

Signature CDB 8/16/10

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 23RD day of SEPT, 2010.

(Notary Seal Stamp)

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

STATE OF ALASKA SCIENTIFIC CRIME DE

DANIEL L. NICHOLAS Commission # 116780 NOTARY PUBLIC My Commission Expires With Office STATE OF ALASKA

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130297

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130297

AUGUST 03, 2010 COB 8/10/10

OPERATOR'S NAME:
HOELSCHER/JAMES/C
OPERATOR'S NUMBER: 5343
SUBJECT'S LAST NAME:
VERIFICATION
SUBJECT'S FIRST NAME/MI :
OF CALIBRATION
O.L. #: N/A
DEPT/AGENCY: HPB1
CASE/REPORT: N/A
TEST TYPE: V
ALCO TARGET VALUE: .081
ALCO S/N: X172968

— BREATH ANALYSIS —

.081 ADJUSTED FOR 30.32 in

ALCO TARGET	.082	14:19
BLANK TEST	.000	14:20
INTERNAL STANDARD	VERIFIED	14:20
ALCO TV 30.32 in	.079	14:20
BLANK TEST	.000	14:21
SUBJECT SAMPLE COB 8/10/10	.000	14:22
BLANK TEST	.000	14:22
ALCO TV 30.32 in	.079	14:23
BLANK TEST	.000	14:24

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130297

AUGUST 03, 2010
TIME 14:17

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	41c
BAROMETER:	30.32 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~