

VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath test Instrument

State of Alaska

Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N: <u>130298</u> ✓
Name: <u>Peter C Steen</u>	ID#: <u>4572</u>	Date: <u>1/28/09</u>
Agency: <u>Palmer Police Department</u>	Phone #: <u>(907) 745-4811</u>	
Instrument Location: <u>Palmer Police Department</u> ✓		
Alco S/N: ^{78268 cos ✓} <u>78259</u> Target Value: ^{077 cos ✓} <u>080</u> High Pressure: <u>1100</u>		
B Alco Test Value Average: ^{078 ✓} <u>078</u> ^{077 ✓} <u>077</u>		
1 st Alco ^{2.10.09} <u>2.10.09</u> 2 nd Alco		
Signature: 		<u>cos</u> <u>2/17/09</u>
(OVER)		



Under the Alaska Rules of Evidence, (Do Not write in the area below) I certify that,

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Peter C Steen, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

02/18/09

Date

VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath test Instrument

State of Alaska

Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program

DataMaster cdm S/N: 130298

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130298

JANUARY 28, 2009 ✓

OPERATOR'S NAME:

STEEN, PETER, C

OPERATOR'S NUMBER: 4572

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI:

VOC

O.L. #: VOC

DEPT/AGENCY: PLMI

CASE/REPORT: VOC

TEST TYPE: U

ALCO TARGET VALUE: .077 ✓

ALCO S/N: 78268

— BREATH ANALYSIS —

.077 ADJUSTED FOR 29.81 in
ALCO TARGET .076 02:59
BLANK TEST .000 03:00
INTERNAL STANDARD VERIFIED 03:00
ALCO TV 29.81 in .078 ✓ 03:00
BLANK TEST .000 03:01
SUBJECT SAMPLE .000 03:01
BLANK TEST .000 03:02
ALCO TV 29.81 in .077 ✓ 03:02
BLANK TEST .000 03:03

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130298

JANUARY 28, 2009 ✓

TIME 02:58

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.81 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmnop
qrs tuvwxyz{|}~■