

Ending VOC 4-11-10

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

APR 13 2010

| | | | |
|------------------------------------------------------------|------------------------------|----------------------|---------------------------------------|
| Supervisor/Operator Performing the Verification Procedure: | | DataMaster cdm S/N | <u>130298</u> ✓ |
| Name | <u>TIM TUCKWOOD</u> | ID# | <u>3446</u> Date <u>4-11-10</u> |
| A Agency | <u>Alaska State Troopers</u> | Phone # | <u>895-4800</u> |
| Instrument Location | <u>Arctic Man Event</u> ✓ | | |
| B Alco S/N | <u>1B</u> ✓ | Target Value | <u>080</u> ✓ High Pressure <u>950</u> |
| Alco Test Values | <u>073</u> ✓ | <u>070</u> ✓ | |
| | 1 st Alco | 2 nd Alco | |
| Signature | <u>Tim Tuckwood</u> ✓ | | |

(OVER)

BMB
4/15/10

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 20th day of April, 2010.

Carolyn M. Noland (Notary Seal Stamp)
 Carolyn M. Noland
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130298

(CONTINUED FROM FRONT PAGE)

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

*Ending VOC
4-11-10*

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130298

APRIL 11, 2010

OPERATOR'S NAME: *AMB 4/15/10*
TUCKWOOD TIM
OPERATOR'S NUMBER: 3446
SUBJECT'S LAST NAME:
VERIF
SUBJECT'S FIRST NAME/MI :
OF CAL
O.L.N. #: 12345
DEPT/AGENCY: DJTB
CASE/REPORT: 09876
TEST TYPE: U
ALCO TARGET VALUE: .088 ✓
ALCO S/N: 1B ✓

--- BREATH ANALYSIS ---

| | | |
|----------------------------|----------|-------|
| .088 ADJUSTED FOR 27.05 in | | |
| ALCO TARGET | .072 | 07:52 |
| BLANK TEST | .000 | 07:52 |
| INTERNAL STANDARD | VERIFIED | 07:52 |
| ALCO TV 27.05 in | .073 | 07:53 |
| BLANK TEST | .000 | 07:54 |
| SUBJECT SAMPLE | .000 | 07:54 |
| BLANK TEST | .000 | 07:55 |
| ALCO TV 27.05 in | .070 | 07:55 |
| BLANK TEST | .000 | 07:56 |

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130298

APRIL 11, 2010
TIME 07:50

--- DIAGNOSTIC CHECK ---

| | |
|------------------|----------|
| COMPUTER: | OKAY |
| PROGRAM: | OKAY |
| SOFTWARE DATE: | 01/08/09 |
| HEATERS | |
| SAMPLE CHAMBER: | 50c |
| BREATH TUBE: | 41c |
| BAROMETER: | 27.05 in |
| FLOW DETECTOR: | OKAY |
| PUMP | |
| HIGH SPEED: | OKAY |
| DETECTOR: | OKAY |
| FILTERS: | OKAY |
| QUARTZ STANDARD: | OKAY |
| CALIBRATION: | OKAY |

PRINTER TEST

!*"#&'()*+,-./0123456789:;<=>?@ABCDEFGHI
HIJKLNMOPQRSTUVWXYZ\]^_`abcde fghi jklmno
pqrstuvwxyz{|}~